

CDC 2019-nCoV Real-Time RT-PCR Diagnostic Panel

Acceptable Alternative Primer and Probe Sets

The following lots of N1, N2, and RP primers and probes have passed functional testing at CDC and may be used with the CDC 2019-nCoV Real-Time RT-PCR Diagnostic Panel Instructions for Use under CDC's Emergency Use Authorization (EUA):

<https://www.fda.gov/medical-devices/emergency-situations-medical-devices/emergency-use-authorizations>

Please note that only the listed lot numbers of material are acceptable for testing under CDC's EUA.

Any primer and probe reagents included in these kits in addition to N1, N2 and RP have not been tested by CDC and may not be used for diagnostic testing under CDC's EUA. Only N1, N2 and RP primer and probe sets may be used.

Manufacturer	Product name	Catalog No.	Acceptable Lots
BioSearch Technologies www.biosearchtech.com 1-800-GENOME-1	2019-nCoV CDC Probe and Primer Kit for SARS-CoV-2	KIT-nCoV-PP1-1000	143503 143764 144000 144604
Integrated DNA Technologies (IDT) www.idtdna.com (800) 328-2661	2019-nCoV Kit, 500 rxn	10006606	0000500383 0000504847 0000505594 0000505969 0000507295 0000507509 0000508150 0000508785 0000509402 0000510344 0000510908 0000511393 0000512209 0000512692 0000513132
Integrated DNA Technologies (IDT) www.idtdna.com (800) 328-2661	2019-nCoV Kit, 1000 rxn	10006770	0000515825 0000540655 0000522562 0000542159 0000524162 0000550657 0000527162 0000553336 0000529986 0000558003 0000536597 0000563404 0000538650 0000566435 0000571931 0000572933 0000573777 0000577471

Last Updated: January 21, 2021

For current list please visit: <https://www.cdc.gov/coronavirus/2019-ncov/lab/virus-requests.html>