

FR Doc #04-7984

PUBLIC COMMENT 8400021

From: "Gregory Hobbs" <ghobbs@microbac.com>
To: <wvogl@samhsa.gov>
Date: 5/20/2004 11:20 AM
Subject: comment on 04-7984

The requirement for confirmation of specific gravity measurements are for refractometry, but require a digital refractometer with a printing capability. This is excessive, since the number of specimens being measured in this way will be relatively small and can be handled manually in an accurate manner by using manual refractometry.

Gregory A. Hobbs, Ph.D.
Director of Toxicology
Microbac Laboratories, Inc.
Louisville, KY 40213
502-962-6400 (Tel)
502-962-6411 (FAX)

This electronic message contains information from Microbac Laboratories, Inc. that may be privileged or confidential. If you have received this message in error, please delete it without copying or forwarding. Please notify the sender. Thank you.

Scanned by WinProxy
<http://www.Ositis.com/>