

**Analysis of the June 8, 1998 Performance Evaluation  
HIV-1 RNA Determinations (Viral Load) Results  
Reported to the Centers for Disease Control and Prevention  
by Laboratories Participating in the Model Performance Evaluation Program**

This report is an analysis of results reported to the Centers for Disease Control and Prevention (CDC) by laboratories participating in the Model Performance Evaluation Program (MPEP) after they performed ribonucleic acid (RNA) determinations on human immunodeficiency virus type 1 (HIV-1) performance evaluation samples shipped to them June 8, 1998. Testing results were reported by 168 (88%) of the 191 laboratories who were sent sample panels.

Samples used in the MPEP HIV-1 RNA determinations performance evaluation survey are undiluted, unpooled plasma obtained from individual donors who are HIV-1 infected or uninfected. Before shipment, the CDC tested each donor with at least three test kits which included the viral RNA test kit approved by the Food and Drug Administration (FDA), and two test kits not approved by the FDA and designated for research use only.

The second page following the report title page, Table 1, lists the CDC panels for this shipment, the labeled vials contained in each panel, the CDC donor numbers, the CDC results obtained with each test kit manufacturer, and the CDC interpretation of the results based on the manufacturers' criteria. For all the HIV-1 infected donors, HIV-1 RNA was detected by all the test kits used and the CDC interpretation for these donors was positive for RNA. Conversely, the donors not infected with HIV-1 did not have HIV-1 RNA detected consistent with the criteria contained within the test kit manufacturer's insert. Based upon the lower limits of the test kit sensitivities, these donors were interpreted by CDC as negative for HIV-1 RNA.

### **Summary of Results**

Figure 1 shows the cumulative frequency of test results reported by laboratories for donors who were HIV-1 infected and had detectable HIV-1 RNA, and for donors not infected with HIV-1 and in whose donor plasma HIV-1 RNA was not detectable. For the samples obtained from donors (Donor 1, Donor 1 duplicate, Donor 2, and Donor 2 duplicate) that were infected with HIV-1, 508 (100%) of the results reported by the participant laboratories indicated HIV-1 RNA was detected. Of the 346 results reported for the samples obtained from donors not infected with HIV-1 (Donor 4 and Donor 5), laboratories reported 343 (99.1%) results that indicated not detecting HIV-1 RNA, while 3 laboratories reported 3 (0.9%) results that indicated detecting HIV-1 RNA. These 3 results were values that were above lower limit sensitivities of the test kits used by these laboratories.

### **Types of Laboratories Performing HIV-1 RNA Determinations**

The types of laboratories reporting results are shown in Figure 2. Each laboratory type is listed by decreasing frequency. Consistent with previous panel surveys, more than 50% of the laboratories that reported results are hospital laboratories.

## **Types of Test Kits Used by Laboratories**

The types of test kits used by laboratories performing viral RNA determinations are shown in Figure 3 and are listed by decreasing frequency. The Roche Amplicor HIV-1 Monitor™ test kit, approved by the FDA, was used by 69% of the laboratories reporting results.

## **Aggregate Testing Results Reported by Donor**

Aggregate testing results, for each donor by test kit, reported by participant laboratories, are shown in Table 2. Since the lower limit sensitivities of the reported test kits ranged from <20 RNA copies/ml to <500 RNA copies/ml, the results are shown for each individual donor by test kit and listed according to the minimum, maximum, and median values that were calculated from the reported results. Information listed in the results section for each individual donor also includes the HIV-1 infection status of the donor and which panel vials contained the donor material. The first page of Table 2 shows the laboratory test results reported for CDC Donor 1 and Donor 1 duplicate, an HIV-1 infected donor. The second page shows the results reported for Donor 2 and Donor 2 duplicate, an HIV-1 infected donor. The third page shows the results reported for Donor 3, an HIV-1 infected donor and Donor 4, an uninfected donor. The fourth page shows the results reported for Donor 5, an uninfected donor. For this shipment, Donor 1 and Donor 2 were samples that were duplicated in each panel providing participant laboratories the opportunity to review their intra-shipment reproducibility for those donor samples.

Please note that in Table 2, the columns under each donor sample provide the number of laboratory results detecting HIV-1 RNA or not detecting HIV-1 RNA, followed by the minimum, median, and maximum result value listed for each test kit manufacturer.

In general, laboratories performed very well in testing these performance evaluation samples. All laboratories detected HIV-1 RNA in those samples obtained from donors infected with HIV-1 and in which CDC detected HIV-1 RNA. There were 11 laboratories that reported invalid testing results for Donor 3 that was associated with unknown inhibition in testing their samples. All of the laboratories were using test kits manufactured by Roche, although no specific kit lot number could be associated with these inhibitory reports.

Similarly, most laboratories did not detect viral RNA in the samples obtained from the donors not infected with HIV-1 (Donor 4 and Donor 5). Two incorrect determinations were reported for donor 4 by laboratories using a Roche Amplicor HIV-1 Monitor™, 10,497 copies/ml, and a Chiron HIV-1 Quantiplex™ test kit, 803 copies/ml. One incorrect determination was reported for Donor 5 by a laboratory using a Chiron HIV-1 Quantiplex™ test kit, 628 copies/ml.

## **Use of Quality Control Testing Material**

Information was collected on the use of quality control (QC) samples in addition to the controls contained in the test kits. Depending on the manufactured test kit used, positive and negative test controls, test standards, or test calibrators are internal kit control samples used to validate a test run and to quantitate HIV-1 RNA copies/ml, and may not validate the analytic testing process

which may include testing problems related to pipetting, inadequate incubation conditions, inadequate washing, or variability in kit lot sensitivity. Of the 168 laboratories that reported results, 164 (97%) laboratories provided information on their use of QC samples other than the controls contained in the test kit. Of these, 52 (32%) indicated they used QC samples other than those contained in the test kit and 51 laboratories indicated their source of QC material. Among these 51 laboratories, 34 (66%) indicated they obtained their QC material from an in house source and 17 (34%) obtained their QC material from a commercial source. Although, some laboratories indicated using a single serum/plasma, or multiple serum/plasma, 4 laboratories indicated they used VQA standards obtained through AIDS Clinical Trial Group (ACTG) participation. Although various combinations of QC materials were used, e.g., high RNA copies plus a negative control or low RNA copies plus a negative control, some laboratories indicated they used a high RNA copy control, low RNA copy control, and negative control all in combination. Of the 51 laboratories providing source and frequency of use information for QC material in addition to that contained in their test kit, 24 (47%) used their QC material with each set of tests, 14 (27%) used QC material only with each new test kit, 8 (16%) used QC material with each new test, 6 (12%) indicated an “Other” use frequency, and one (2%) indicated they used QC material daily.

## **Conclusion**

The results of this third performance evaluation shipment for HIV-1 RNA determinations showed that all laboratories correctly detected HIV-1 RNA in those samples from donors infected with HIV-1. Similarly, most laboratories did not detect HIV-1 RNA in the samples from donors not infected with HIV-1, while only three laboratories did detect HIV-1 RNA in these donor samples. While there is variability of results within a kit manufacturer and between kit manufacturers, a comparison of the results reported for the duplicate donors for this performance survey shipment showed that most results were reproducible. For the samples from donors infected with HIV-1, the overall analytic sensitivity for the results reported was 100%. For the samples from donors not infected with HIV-1, the overall analytic specificity was 99.1%.